## Pixel Failure Modes - Local

DOCDB 404-v1

PIXEL FAILURE MODES Lucien Cremaldi


06 Sep 2005

Problem (Local)	Siemens	FEC/DCU	FED	Chiller Plant	CAEN	GC/MS
	T / %H	Temp	Temp/Rate	Flow/Wght	ΔI-Current	%Gas
6-U cooling loop failure (small leak )	ΔT<2	ΔT<2	ΔT<2	0	0	%FC
6-U cooling loop failure (large leak/break/blocked)	ΔT>50	ΔT>50	ΔT>50	0	0	%FC
Supply line failure (small leak)	ΔT<2	ΔT<2	ΔT<2	0	0	%FC
Supply line failure (large leak/break/blocked)	ΔT>50	ΔT>50	ΔT>50	FLOW	ΔΙ	%FC
Poor contact between frame and pipe.	∆T<2	∆T<2	ΔT<2	0	0	
%Humidity increase	%Н	0	0	0	0	%H2O
Local HV/LV Dead Short	ΔΤ	ΔΤ	ΔΤ	0	ΔΙ	0
Local HV/LV Leakage Current High	ΔΤ	ΔΤ	ΔΤ	0	ΔΙ	
Excessive Plaquette Vibration	Accelerom eter					
Noise/Grounding Loop			RATE		$\Delta I$	


## Pixel Failure Modes - Global

Problem (Global)	Siemens	FEC/DCU	FED	Chiller		GC/MS
	T / %H	Temp	Temp/Rate	Flow/Wght	ΔI-Current	%Gas
Si Tracker - Thermal screen Off or Failure	ΔT<2	ΔT<2	ΔT<2	FLOW	ΔΙ	0
Cooling Plant global failure (stop, major leak)	∆T>50	∆T>50	∆T>50	FLOW	ΔΙ	0
Major leak in supply lines.	∆T>50	∆T>50	ΔT>50	FLOW WGHT	ΔΙ	0
Minor leak in supply lines.	0	0	0	FLOW	0	0
High radiation level	DOSE	DOSE	0	0	ΔΙ	0
N2 flow failure (stop flowing, no dry)	0	0	0	0	0	%H2O
Global power cut in the counting room						0
Global power cut in the cavern	ΔΤ	ΔΤ	0	0	ΔΙ	0
Siemens PLC failure	PLC	0	0	0	0	0
Coolant radiation damage or impurity pollution.	ΔΤ	ΔΤ	ΔΤ	FC72	ΔΙ	0

## Interface to DCS



- Control Scripts written by software experts.
- Important input from hardware: TB and R&D experience
- •Timing Sequence defined for each.



## **Engineering Tests**

